



# Penn

## CAD/CAM-Ceramic-Center

Director: Michael Bergler, MDT  
School of Dental Medicine,  
Robert Schattner Building  
240 S. 40<sup>th</sup> Street,  
Philadelphia, PA 19104  
T. 215-746-2075 F. 215-898-9981  
e-mail. [mbergler@dental.upenn.edu](mailto:mbergler@dental.upenn.edu)

### Flexural Strength Test Result of Anterior Zirconia, Zirkozahn GmbH, Gais, Italy

Material	Flexural Strength MPa
ANTERIOR TO	670

Dr. Francis K. Mante  
University of Pennsylvania  
School of Dental Medicine

Michael Bergler, CDT, MDT  
University of Pennsylvania  
School of Dental Medicine